



# Certificate of Traceability

## Level I Laboratory Certification

This certifies that **ICON Central Laboratory - India**, using **Bio-Rad Variant II** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial Reference** method. The comparison was performed with: **University of Missouri SRL#7**.

The system evaluated was:

Instrument: <b>Variant II</b>	Calibrator Lot: <b>82114, 82115</b>	Column Lot: <b>NU62000HR2</b>
Reagent Lot: <b>AA82340NU, AA82341NU, AA82342NU</b>	Calibrator Assigned Values: <b>5.4%, 13.0%</b>	

Date of Certification: May 1, 2009

Certification Expires: May 1, 2010

*David South*

NGSP Steering Committee Chair

*Randie R Little PhD.*

NGSP Network Coordinator

*Sharon Amulfo*

SRL director/ supervisor